

WPS500X

500 psi / 34.5 bar

Automotive Pressure Transducer

This accessory complies with the following standards:

- Electromagnetic Compatibility (EMC) Directive 2004/108/EC
- FCC Part 15
- Pressure Equipment Directive 97/23/EC Article 3.3



1. Safety terms and symbols

Appearing in this manual:

\triangle	WARNING statements identify conditions or practices that could result in injury or death.
WARNING	
CAUTION	CAUTION statements identify conditions or practices that could result in damage to this product or other property.

Appearing on the product:



Danger of personal injury or property damage. Refer to manual for details.

General safety summary

Please review the following safety precautions to avoid injury and prevent damage to this transducer or any equipment that is connected to it.

DO connect safely

This transducer must be connected using only the pressure hoses and connectors supplied. Pico Technology cannot accept responsibility for damage or injury caused by the use of unsuitable pressure hoses or connectors.

DO follow the vehicle manufacturer's safety instructions

This is particularly important when connecting the transducer to pressurized fuel lines.

DO ground the unit when working with fuels

When working on or near fuel systems, you must connect the transducer to an electrical ground. If you are using the transducer with a PicoScope oscilloscope, connect a suitable lead from any unused BNC connector on the front panel of the scope to the ground of the vehicle.

DO wear appropriate personal protective equipment

Wear appropriate personal protective equipment (PPE) when working with pressurized fluids.

DO NOT exceed maximum working pressure

To avoid injury, do not use the transducer with pressures above 500 psi / 34.5 bar.

DO NOT exceed maximum temperature

The lithium polymer (LiPo) battery inside the transducer can be damaged by excessive heating. Do not store or operate the unit at temperatures above 60 °C (140 °F).

DO NOT operate without covers

To avoid equipment damage and personal injury, do not operate this transducer with the covers removed.

DO NOT operate in wet or damp conditions

To avoid incorrect readings and possible equipment damage, do not operate this transducer in wet or damp conditions or submerge it in liquid. The transducer is splash-resistant but not immersion-proof.

DO NOT operate in an explosive atmosphere

To avoid personal injury and fire hazard, do not operate this transducer in an explosive atmosphere.

DO NOT operate the transducer if damaged

If you suspect there is damage to this transducer, have it inspected by qualified service personnel. Do not attempt to dismantle or repair it yourself.

DO NOT use with damaged pressure hoses

All the pressure hoses supplied by Pico Technology have been pressure-tested and must not be used if they have been disassembled or damaged.

DO NOT use in long-term or permanent installations

The WPS500X is intended for immediate diagnostic purposes and not long-term or permanent use. For example, it should not be connected to a race car as a monitoring system.

DO NOT use on high pressure common-rail diesel fuel systems

The high pressure from the mechanical fuel pump (usually found in the engine bay) cannot be measured using the WPS500X. However, the pressure from the electronic fuel pump (usually found close to the fuel tank) can be measured using the WPS500X.

2. Description

The WPS500X automotive diagnostic pressure transducer allows quick and accurate pressure analysis of many automotive systems. It can be used for many different pressure diagnostic applications, saving the need to own several transducers for several applications.

It offers these features:

- high resolution and accuracy
- auto-zeroing
- built-in zoom tool
- integrated bleed-off / pressure relief valve
- three pressure ranges

3. Re-ordering codes

If you need to re-order spare parts, please use the part numbers listed here.

WPS500X pressure transducer (PP652)

Part No	Qty	Description
TA071	1	WPS500X pressure transducer
TA081	1	USB to mini-USB charging cable
TA098	1	5 m (16-foot) BNC to BNC cable
DO157	1	WPS500X manual
MI220	1	Hard plastic carrying case

WPS500X pressure transducer kit (PP836)

Part No	Qty	Description
TA071	1	WPS500X pressure transducer
TA081	1	USB to mini-USB charging cable
TA083	1	Gasoline pressure fuel hose with large Schrader valve
TA085	1	Vacuum hose
TA086	1	Bleed hose
TA087	1	Exhaust adaptor
TA088	1	M14 high-strength compression hose with adapter for deep
		reach spark plugs
TA098	1	5 m (16-foot) BNC to BNC cable
TA103	1	M10 high-strength compression hose
TA105	1	Ford Triton compression hose
TA117	1	Gasoline pressure fuel hose with small Schrader valve
TA129	1	Universal vacuum adaptor
DO157	1	WPS500X manual
MI220	1	Hard plastic carrying case

Optional accessories

The following accessories are also available for purchase:

Part No	Description
TA104	M18 high-strength compression hose
TA142	Foster 2 Series quick coupler female to 1/8" MPT male for making custom hoses

O-rings

The following O-rings can be used as replacements on our compression hoses.:

Used on Compression Hose M10 (TA103)	OD 7/16″ ID 5/16″	1/16″	Viton
Used on Compression Hose M14 (TA088)	OD 9/16″ ID 7/16″	1/16″	Viton
Used on Compression Hose M18 (TA104)	OD 11/16" ID 9/16"	1/16″	Viton
ι (ι (TA103) Jsed on Compression Hose M14 TA088) Jsed on Compression Hose M18 TA104)	TA103)ID 5/16"Jsed on Compression Hose M14OD 9/16"TA088)ID 7/16"Jsed on Compression Hose M18OD 11/16"Jsed on Compression Hose M18ID 9/16"	ID 5/16" I/16" Jsed on Compression Hose M14 OD 9/16" Jsed on Compression Hose M14 OD 9/16" Jsed on Compression Hose M18 OD 11/16" Jsed on Compression Hose M18 OD 11/16" Jsed on Compression Hose M18 ID 9/16"

*Pico Technology does not stock these parts.

4. The parts of your WPS500X pressure transducer



Signal output	Use the BNC-to-BNC cable supplied to connect this to your
	oscilloscope.
Front-panel controls	See Section 5.
USB charging port	For battery recharging only (no data connection). Connect
	to any USB port on a computer or a 5 volt USB wall charger.
Inlet port	Connect the pressure hose here.
Bleed vent	When the bleed screw is opened, this vent allows fluid to drain out of the measurement chamber.
Bleed button	Push to open the bleed vent. Some models have a screw instead of a push-button.

5. Front-panel controls



6. Operating modes

Ranges 1 to 3

Select the range using the Range button shown in Section 5: "Front-panel controls".

Range 1: The first range measures from -15 psi (-1 bar) to +500 psi (34.5 bar). It offers high resolution and accuracy for high-pressure tests such as cranking and running cylinder compression or fuel pressure testing. It is also helpful for identifying cam timing issues such as jumped timing belts and stretch timing chains, especially on multi-cam engines that may not have a cam sensor on each camshaft. The WPS500X kit includes a specially designed compression hose for performing this test with a significantly reduced error and higher operating temperature range than conventional compression hoses and adapters.

Range 2: The second range measures from -15 psi (-1 bar) to +50 psi (+3.45 bar). This range is ideal for vacuum test and fuel system tests. When testing these systems the zoom function is especially useful to analyze valve operation with the vacuum waveform or the injectors through the fuel waveform.

Range 3: The third range measures -5 to +5 psi (-0.345 to +0.345 bar). This setting is sensitive enough to analyze small pressures or pulses such as exhaust pulses from the tail pipe.

Zoom modes 1 to 3

The zoom feature operates by removing all of the voltage from the signal below 100 Hz and then magnifying the remaining signal. Select the zoom using the zoom button shown in Section 5: "Front-panel controls".

Zoom 1: This mode multiplies the signal 10 times.

Zoom 2: This mode multiplies the signal 100 times. For example, let's say the transducer is connected to intake manifold vacuum on a running engine at idle with the transducer set to mode 2 (-15 to +50 psi or -1 to +3.45 bar) and the waveform on the scope is at -9 psi (-0.6 bar). The fluctuation in the signal caused by the valves' opening and closing compared with atmospheric pressure forcing air into the manifold will be visible but will only make up a small portion of the overall signal. Selecting the zoom function will bring the waveform up to the zero line (remove the -9 psi portion) and the fluctuations will be magnified, making analysis of this component of the signal much clearer.

Zoom 3: Zoom level 3 multiplies the signal by about 1,000 times. It is extremely sensitive and is suitable only for vacuum measurements.

7. Preparation for use

Compatible fluid types

The WPS500X is suitable for use with the following types of fluid:

- Low-pressure gasoline and diesel
- Engine oil
- Air

Before first use

- Remove all packaging.
- Charge the internal battery. For instructions, see Section 9: "Maintenance".

8. Making a measurement

Making a pressure measurement involves the following steps. Each step is explained in more detail in the sections below.

- Prepare the transducer
- Measure pressure
- Release the pressure in the measurement chamber
- Clean the measurement chamber

Preparing the transducer

- Ensure that the transducer's internal battery is charged.
- Unplug the charging cable from the transducer.
- Before switching on, disconnect any pressure source from the transducer. A
 pressure source left connected will interfere with the self-calibration procedure
 described below.
- Switch on the transducer and wait until the three Range LEDs light up in sequence. At the end of the sequence, Range 1 LED remains lit, showing that the transducer has finished its auto-zeroing procedure. The entire procedure should take less than 10 seconds.

Measuring pressure

- If your WPS500X has a bleed screw, ensure that the bleed screw is firmly closed.
- Connect the appropriate pressure hose to the pressure sensing port.



Use only pressure hoses and connectors supplied with the WPS500X or otherwise manufactured to an adequate standard. Pico Technology cannot accept responsibility for damage or injury caused by the use of unsuitable pressure hoses or connectors.



Always check that the pressure hose is securely fastened to the transducer before pressurizing the system. Always check for leaks when connecting the unit and to never leave it connected to a vehicle unattended (especially when connected to fuel).

- If measuring liquid pressure, bleed any air out of the measurement chamber first.
- Use the BNC cable supplied to connect the output of the transducer to the input channel of the oscilloscope.
- Switch on the computer and run the PicoScope software.
- In the PicoScope software, select the "Automotive" menu and then the appropriate pressure test.
- Press the Range button on the transducer to select the desired measuring range.
- Start the vehicle's engine.
- A waveform showing the pressure of the system will appear on the PicoScope display.

Releasing the pressure in the measurement chamber

After each measurement, some fluid will remain under pressure in the measurement chamber. Follow the instructions below to release the pressure.

• Hold the transducer over a suitable container to catch the fluid expelled from the bleed vent.



The fluid released from the bleed vent may be under high pressure. Position the transducer so that the fluid released cannot cause damage or injury.

Push-button units

Press the button to release the pressure from the chamber.

Bleed-screw units

- **SLOWLY** loosen the bleed screw by turning it counter-clockwise. Do not remove the screw from the transducer.
- Allow the fluid to emerge from the bleed vent.
- When no more fluid emerges, tighten the bleed screw.

Cleaning the measurement chamber

If you are measuring the pressure of a liquid, some of the liquid will remain in the measurement chamber after use. To prevent cross-contamination between liquids, or between liquids and air, you must clean the measurement chamber after use.



DISCONNECT all pressure hoses from the transducer. **DO NOT** attempt to clean the measurement chamber when the unit is under pressure.

When disassembling the bleed valve for the first time, hold the transducer over an empty container in case any loose parts drop out.

• Dismantle the bleed valve and allow any liquid to drain out.

The bleed valve in your WPS500X pressure transducer may be either a ball valve, a needle valve or a Schrader valve, as shown in the drawings below.



WPS500X with ball valve



WPS500X with needle valve



WPS500X with push-button valve

Ball valve

- If the ball is still inside the ball valve, allow it to drop out. Clean and oil the ball with a non-hygroscopic oil that is compatible with steel and aluminum. Ordinary motor oil is suitable.
- If the ball needs replacing, see Section 9: "Maintenance".
- Clean the bleed screw. Flush out the measurement chamber with brake cleaner if necessary, then replace the ball and the bleed screw.

Needle valve

• Clean the bleed screw. Flush out the measurement chamber with brake cleaner if necessary, then replace the bleed screw.

Push-button valve

- Remove the countersunk screw from the back of the transducer near the push-button.
- Hold the transducer upside-down with the screw hole facing down.
- Prepare to catch the retaining pin that will drop out of the hole in the next step.
- Depress the push-button until the retaining pin drops out.
- Withdraw the push-button assembly, including the spring.
- Clean the valve parts, spring, retaining pin and screw with brake cleaner.
- Flush out the measurement chamber with brake cleaner. Push in the centre of the Schrader valve, visible at the bottom of the push-button hole, to allow the cleaner to flow through.
- Replace the push-button, retaining pin and screw. When replacing the screw, take care not to allow any grit into the mechanism. The screw should be hand-tight.

9. Maintenance

Cleaning the housing

Clean the transducer's housing by wiping it with a rag moistened with clean water or water-based detergent. Allow the housing to dry before use.

- Do not use fuel or any other solvent
- Do not use abrasive cleaning agents
- Do not submerge the unit in any liquid
- Do not dismantle the unit
- Do not use the unit until it is perfectly dry.

Cleaning the measurement chamber

See Section 8 above.

Recharging the internal battery

- To charge using a computer, switch on the computer and allow it to boot. Disable any power-saving modes to ensure that the computer does not switch off before recharging is complete. Connect the transducer to the USB port of the computer using the USB charging cable provided.
- To charge using a USB wall charger, connect the transducer to the charger using the USB charging cable provided.
- Leave the transducer to charge for 5 hours.
- Unplug the USB cable from the transducer before use. Leaving the cable plugged in may affect measurement accuracy.

Repairs

If the unit is damaged or stops working, return it to Pico Technology or an authorized Pico distributor for repair. Do not attempt to dismantle or repair the unit.

Replacing the bleed valve ball (if fitted)

If the ball in the bleed valve is lost, it can be replaced with a standard 0.25 inch (6.35 mm) steel bearing ball. To re-seat the ball, clean the port and then seat the ball using a brass drift.

Disposal

The WPS500X contains a lithium polymer (LiPo) battery. When the transducer reaches the end of its life, take the entire unit to a battery recycling facility.



YOU MUST observe the instructions below. Incorrect disposal of the battery could cause a fire or an explosion.

- Do not open the unit to remove the battery
- Do not crush or shred the unit
- Do not dispose of in fire

10. Specifications

Inlet			
Pressure ranges	Range 1	Range 2	Range 3
	-15 to +500 psi -1 to +34.5 bar	-15 to +50 psi -1 to +3.45 bar	-5 to +5 psi -0.345 to +0.345 bar
Connector		Male push-fit	
Output			
Scaling	1 V/100 psi (6.89 bar)	1 V/10 psi (0.689 bar)	1 V/1 psi (0.0689 bar)
Offset (typical)		Auto-zeroing	
Connector	BNC female, f	its Pico Technolog	gy cable TA098
Performance			
Accuracy	1% of scale	1% of scale	5% of scale
Response time (10% to 90%)	100 µs	100 µs	filtered
Power supply			
Туре	Built-in LiPo	battery, not use	r-serviceable
Charging current	500 mA froi	(max.) at 4.75 V m USB charger c	to 5.25 V able
Charging connector	USB mini, fit	s Pico Technolog	y cable TA081
Ambient operating temperature	0 to 60	°C (32 to 140 °	F) max.
Ambient storage temperature	0 to 60	°C (32 to 140 °	F) max.
Environmental protection	nvironmental protection Splash-resistant against water, gasoline a diesel. Not immersion-proof.		r, gasoline and -proof.
Weight		332 g (11.7 oz)	
Dimensions	133 x 74 x	30 mm (5.2 x 2	.9 x 1.2 in.)

11. Conversion factors

The SI unit of pressure and vacuum is the pascal, symbol Pa. These are some other units in common use:

1 bar	100 000 Pa
1 psi (pound per square inch)	≈ 6 895 Pa
1 inHg (inch of mercury)	≈ 3 386 Pa
1 inH ₂ O (inch of water)	≈ 250 Pa
1 mmH ₂ O (millimeter of water)	≈ 10 Pa

Issue	history:
-------	----------

1)	1.6.09	First issue.
2)	14.7.09	Corrections to power LED description, etc.
3)	21.9.09	Added FCC statement, needle-valve version.
4)	26.4.10	Removed kit. Added accessories list. Mentioned brake
		cleaner for flushing.
5)	6.7.10	Added push-button valve, temperature lower limits.
6)	17.2.12	-Refs to hydraulic fluids +PP836 contents list
		+Low-pressure fuel only.
7)	14.05.12	O rings information added.

Pico Technology James House Colmworth Business Park St. Neots PE19 8YP United Kingdom

www.picoauto.com

This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

Pico Technology is a registered trademark of Pico Technology Ltd.

CE





Copyright © Pico Technology Ltd. 2009-2012