



PicoLog 1000 Series

Programmer's Guide

Contents

1 Introduction	1
1 Overview	1
2 Legal information	2
3 Contact details	3
2 Getting started	4
1 About the driver	4
2 Installing the driver	4
3 Connecting the logger	4
4 USB ADC-11 compatibility mode	5
3 Technical reference	6
1 Capture modes	6
2 Scaling	6
3 Driver routines	7
1 Summary	7
2 pl1000CloseUnit	8
3 pl1000GetSingle	9
4 pl1000GetUnitInfo	10
5 pl1000GetValues	11
6 pl1000MaxValue	12
7 pl1000OpenUnit	13
8 pl1000OpenUnitAsync	14
9 pl1000OpenUnitProgress	15
10 pl1000Ready	16
11 pl1000Run	17
12 pl1000SetDo	18
13 pl1000SetInterval	19
14 pl1000SetPulseWidth	20
15 pl1000SetTrigger	21
16 pl1000Stop	22
17 PICO_STATUS values	23
4 Example programs	25
1 Introduction	25
2 C and C++	25
3 Delphi	25
4 Excel	25
5 LabVIEW	26
6 Visual Basic	26
7 Agilent-VEE	26
4 Glossary	27
Index.....	29

1 Introduction

1.1 Overview

The PicoLog 1000 Series PC Data Loggers are medium-speed, multichannel voltage-input devices for sampling analog data using a PC. This manual explains how to use the Application Programming Interface and drivers to write your own programs to control the unit. You should read it in conjunction with the [PicoLog 1000 Series User's Guide](#).



The following PicoLog 1000 Series Data Loggers are available:

Version	Part No.	Resolution	Channels
PicoLog 1012	PP543	10 bits	12
PicoLog 1216	PP544	12 bits	16

These devices can also be used with the PicoLog data logging software and the PicoScope oscilloscope software.

1.2 Legal information

The material contained in this release is licensed, not sold. Pico Technology grants a licence to the person who installs this software, subject to the conditions listed below.

Access

The licensee agrees to allow access to this software only to persons who have been informed of these conditions and agree to abide by them.

Usage

The software in this release is for use only with Pico products or with data collected using Pico products.

Copyright

Pico Technology claims the copyright of, and retains the rights to, all material (software, documents etc.) contained in this release. You may copy and distribute the entire release in its original state, but must not copy individual items within the release other than for backup purposes.

Liability

Pico Technology and its agents shall not be liable for any loss, damage or injury, howsoever caused, related to the use of Pico Technology equipment or software, unless excluded by statute.

Fitness for purpose

No two applications are the same: Pico Technology cannot guarantee that its equipment or software is suitable for a given application. It is your responsibility, therefore, to ensure that the product is suitable for your application.

Mission-critical applications

This software is intended for use on a computer that may be running other software products. For this reason, one of the conditions of the licence is that it excludes usage in mission-critical applications, for example life support systems.

Viruses

This software was continuously monitored for viruses during production, but you are responsible for virus-checking the software once it is installed.

Support

If you are dissatisfied with the performance of this software, please contact our technical support staff, who will try to fix the problem within a reasonable time scale. If you are still dissatisfied, please return the product and software to your supplier within 28 days of purchase for a full refund.

Upgrades

We provide upgrades, free of charge, from our web site. We reserve the right to charge for updates or replacements sent out on physical media.

Trademarks

Pico Technology, PicoScope, PicoLog, DrDAQ and EnviroMon are trademarks of Pico Technology Ltd., registered in the United Kingdom and other countries. Pico Technology acknowledges the following product names as trademarks of their respective owners: Windows, Excel, Visual Basic, LabVIEW, Agilent VEE, HP VEE, Delphi.

1.3 Contact details

Address: Pico Technology
James House
Colmworth Business Park
ST. NEOTS
Cambridgeshire
PE19 8YP
United Kingdom

Phone: +44 (0) 1480 396 395
Fax: +44 (0) 1480 396 296

Email:
Technical Support support@picotech.com
Sales sales@picotech.com

Web site: www.picotech.com

2 Getting started

2.1 About the driver

The PicoLog 1000 Series units are supplied with a kernel driver, and a DLL containing routines that you can build into your own programs. The driver is supported by the following operating systems:

- Windows XP (SP2 or later)
- Windows Vista
- Windows 7

Once you have installed the software, the installation directory will contain the drivers and a selection of examples of how to use the drivers. It also contains a copy of this help file in PDF format.

The DLL can be used with any programming language or application that can interface with DLLs: for example, C, Delphi, Visual Basic, Excel and LabVIEW. The [Examples](#) directory contains example programs for several popular programming languages and applications. Some of these examples are fairly simple, but the C console mode example, `pl1000con.c`, illustrates how to use all the facilities available in the driver.

The driver supplied is a 32-bit DLL. However, it will still run on a 64-bit Windows system if you write a 32-bit application and run it under WoW64.

The driver supports up to four USB units at one time.

2.2 Installing the driver

The driver is supplied with the PicoLog 1000 Series SDK. You can download the latest version of the SDK from our website at:

<http://www.picotech.com/software.html>

Select "PicoLog 1000 Series" as the hardware product and then "SDK" as the software product.

The easiest way to install the PicoLog 1000 Series kernel driver is to install PicoScope or PicoLog. These programs are available free of charge from the Pico Technology website. If you prefer to install the driver manually, then proceed as follows:

1. Go to the directory where you unzipped the PicoLog 1000 Series SDK
2. Go to the subdirectory called `system\`
3. Run the program called `dpinst.exe`
4. Plug the device in and follow the instructions in New Hardware Wizard

2.3 Connecting the logger

Before you connect your logger, you must first [install the driver](#) .

To connect the data logger, plug the cable provided into any available USB port on your PC. The first time you connect the unit, Windows will display a New Hardware Wizard. Follow any instructions in the Wizard and wait for the driver to be installed. The unit is then ready for use.

2.4 USB ADC-11 compatibility mode

The PicoLog 1000 Series data loggers may be used as replacements for the USB ADC-11, an 11-channel data logger previously available from Pico Technology. The 1000 Series units have all the functions of the USB ADC-11 and some extra functions, such as extra [digital outputs](#)^[18], a [PWM output](#)^[20] and a sensor power output.

The 1000 Series units are API-compatible with the USB ADC-11. This means that any programs that you have already written do not need to be changed or recompiled - you simply need to update the `usbadc11.dll` to the latest version supplied in the PicoLog 1000 Series SDK. The 1000 Series unit will behave like a USB ADC-11 and the extra outputs (pins 15 to 25) will be internally disconnected. You can continue to use the unit with an old ADC-11 terminal board if you have one, or you can switch to the new Small Terminal Board (PP545).

If you wish to use the extra functions of the 1000 Series units, then you must rewrite your application to use the new PicoLog 1000 Series DLL (`pl1000.dll`), which is described in this manual and is available free of charge from Pico Technology. The SDK is supplied with example code to help you make the transition.

3 Technical reference

3.1 Capture modes

Three modes are available for capturing data:

- `BM_SINGLE`: collect a single block of data and exit
- `BM_WINDOW`: collect a series of overlapping blocks of data
- `BM_STREAM`: collect a continuous stream of data

`BM_SINGLE` is useful when you wish to collect data at high speed for a short period: for example, to collect 1000 readings in 50 milliseconds.

`BM_WINDOW` is useful when collecting several blocks of data at low speeds - for example when collecting 10,000 samples over 10 seconds. Collecting a sequence of single blocks like this would take 10 seconds for each block, so displayed data would not be updated frequently. Using windowing, it is possible to ask for a new block more frequently, for example every second, and to receive a block containing 9 seconds of repeat data and 1 second of new data. The block is effectively a 10-second window that advances one second per cycle.

`BM_STREAM` is useful when you need to collect data continuously for long periods. In principle, it could be used to collect data indefinitely. Every time `pl1000GetValues`^[11] is called, it returns the new readings since the last time it was called. The `noOfValues` argument passed to `pl1000Run`^[17] must be sufficient to ensure that the buffer does not overflow between successive calls to `pl1000GetValues`^[11]. For example, if you call `pl1000GetValues`^[11] every second and you are collecting 500 samples per second, then `noOfValues` must be at least 500, or preferably 1000, to allow for delays in the operating system.

3.2 Scaling

The PicoLog 1000 Series devices produce values in the range 0 to `maxValue`, where `maxValue` is the value returned by the `pl1000MaxValue`^[12] function. To convert ADC readings to volts, multiply by 2.5 and divide by `maxValue`.

For example, `maxValue` for the PicoLog 1216 is 4095. Therefore, an ADC reading of 132 from this device represents $132 \times 2.5 / 4095 = \text{approx. } 0.0806$ volts.

3.3 Driver routines

3.3.1 Summary

The following table explains each of the driver routines supplied with the PicoLog 1000 Series data loggers:

Routine	Description
pl1000CloseUnit ^[8]	close the unit
pl1000GetSingle ^[9]	get a single value from a specified channel
pl1000GetUnitInfo ^[10]	return various items of information about the unit
pl1000GetValues ^[11]	get a number of sample values after a run
pl1000MaxValue ^[12]	return the maximum ADC value
pl1000OpenUnit ^[13]	open and enumerate the unit
pl1000OpenUnitAsync ^[14]	open the unit without waiting for completion
pl1000OpenUnitProgress ^[15]	report progress of pl1000OpenUnitAsync ^[14]
pl1000Ready ^[16]	indicate when pl1000Run ^[17] has captured data
pl1000Run ^[17]	tell the unit to start capturing data
pl1000SetDo ^[18]	control the digital outputs on the unit
pl1000SetInterval ^[19]	set the sampling speed of the unit
pl1000SetPulseWidth ^[20]	configure the PWM output
pl1000SetTrigger ^[21]	set the trigger on the unit
pl1000Stop ^[22]	abort data collection

The driver allows you to do the following:

- Identify and open the logger
- Take a single reading from a particular channel
- Collect a block of samples at fixed time intervals from one or more channels
- Set up a trigger event for a particular channel

You can specify a sampling interval from 1 microsecond to 1 second. The shortest interval that the driver will accept depends on the [capture mode](#)^[6] selected.

The normal calling sequence to collect a block of data is as follows:

```

Check that the driver version is correct
Open the driver
Set trigger mode (if required)
Set sampling mode (channels and time per sample)

While you want to take measurements,
    Run
    While not ready
        Wait
    End while
    ... Get a block of data ...
End While
Close the driver
  
```

3.3.2 pl1000CloseUnit

```
PICO_STATUS pl1000CloseUnit(  
    short handle  
)
```

This function closes the unit.

Arguments:	<code>handle</code> : handle returned from pl1000OpenUnit ¹³ or pl1000OpenUnitProgress ¹⁵
Returns:	PICO_OK ²³ PICO_HANDLE_INVALID ²³

3.3.3 pl1000GetSingle

```

PICO_STATUS pl1000GetSingle(
    short      handle,
    PL1000_INPUTS channel,
    unsigned short * value
)

```

This function returns a single sample value from the specified input channel.

Arguments:	<p><code>handle</code>: handle returned from pl1000OpenUnit^[13] or pl1000OpenUnitProgress^[15]</p> <p><code>channel</code>: which channel to sample, from <code>PL1000_CHANNEL_1</code> to <code>PL1000_CHANNEL_12</code> (for the PicoLog 1012) or <code>PL1000_CHANNEL_16</code> (for the PicoLog 1216)</p> <p><code>value</code>: a location where the function will write the sample value^[6]</p>
Returns:	<p>PICO_OK^[23]</p> <p>PICO_INVALID_HANDLE^[23]</p> <p>PICO_NO_SAMPLES_AVAILABLE^[23]</p> <p>PICO_DEVICE_SAMPLING^[23]</p> <p>PICO_NULL_PARAMETER^[23]</p> <p>PICO_INVALID_PARAMETER^[23]</p> <p>PICO_DATA_NOT_AVAILABLE^[23]</p> <p>PICO_INVALID_CALL^[23]</p> <p>PICO_NOT_RESPONDING^[23]</p> <p>PICO_MEMORY^[23]</p>

3.3.4 pl1000GetUnitInfo

```
PICO_STATUS pl1000GetUnitInfo(
    short      handle,
    char       * string,
    short      stringLength,
    short      * requiredSize,
    PICO_INFO  info
)
```

This function returns a string containing the specified item of information about the unit.

If you want to find out the length of the string before allocating a buffer for it, then call the function with `string = NULL` first.

Arguments:	<p><code>handle</code>: handle returned from pl1000OpenUnit^[13] or pl1000OpenUnitProgress^[15]</p> <p><code>string</code>: a location where the function writes the requested information, or <code>NULL</code> if you are only interested in the value of <code>requiredSize</code></p> <p><code>stringLength</code>: the maximum number of characters that the function should write to <code>string</code></p> <p><code>requiredSize</code>: a location where the function writes the length of the information string before it was truncated to <code>stringLength</code>. If the string was not truncated then <code>requiredSize</code> will be less than or equal to <code>stringLength</code>.</p> <p><code>info</code>: the information that the driver should return. These values are specified in <code>picoStatus.h</code>:</p> <pre>PICO_DRIVER_VERSION PICO_USB_VERSION PICO_HARDWARE_VERSION PICO_VARIANT_INFO PICO_BATCH_AND_SERIAL PICO_CAL_DATE PICO_KERNEL_DRIVER_VERSION</pre>
Returns:	<p>PICO_OK^[23]</p> <p>PICO_INVALID_HANDLE^[23]</p> <p>PICO_NULL_PARAMETER^[23]</p> <p>PICO_INVALID_INFO^[23]</p> <p>PICO_INFO_UNAVAILABLE^[23]</p>

3.3.5 pl1000GetValues

```
PICO_STATUS pl1000GetValues(
    short          handle,
    unsigned short * values,
    unsigned long  * noOfValues,
    unsigned short * overflow,
    unsigned long  * triggerIndex
)
```

This function is used to get values after calling [pl1000_run](#)^[17].

Arguments:	<p>handle: handle returned from pl1000OpenUnit^[13] or pl1000OpenUnitProgress^[15]</p> <p>values: an array of sample values returned by the function. The size of this buffer must be the number of enabled channels multiplied by the number of samples to be collected.</p> <p>noOfValues: on entry, the number of sample values per channel that the function should collect. On exit, the number of samples per channel that were actually written to the buffer.</p> <p>overflow: on exit, a bit field indicating which, if any, input channels overflowed the input range of the device. A bit set to 1 indicates an overflow. The least significant bit corresponds to channel 1. May be <code>NULL</code> if an overflow warning is not required.</p> <p>triggerIndex: on exit, a number indicating when the trigger event occurred. The number is a zero-based index to the <code>values</code> array, or <code>0xFFFFFFFF</code> if the information is not available. On entry, the pointer may be <code>NULL</code> if a trigger index is not required.</p>
Returns:	<p>PICO_OK^[23]</p> <p>PICO_INVALID_HANDLE^[23]</p> <p>PICO_NO_SAMPLES_AVAILABLE^[23]</p> <p>PICO_DEVICE_SAMPLING^[23]</p> <p>PICO_NULL_PARAMETER^[23]</p> <p>PICO_INVALID_PARAMETER^[23]</p> <p>PICO_TOO_MANY_SAMPLES^[23]</p> <p>PICO_DATA_NOT_AVAILABLE^[23]</p> <p>PICO_INVALID_CALL^[23]</p> <p>PICO_NOT_RESPONDING^[23]</p> <p>PICO_MEMORY^[23]</p>

3.3.6 pl1000MaxValue

```
PICO_STATUS pl1000MaxValue(
    short      handle,
    unsigned short * maxValue
)
```

This function returns the maximum ADC value that the device will return. This value may be different on different models in the PicoLog 1000 Series.

Arguments:	<p><code>handle</code>: handle returned from pl1000OpenUnit^[13] or pl1000OpenUnitProgress^[15]</p> <p><code>maxValue</code>: a location where the function will write the maximum ADC value</p>
Returns:	<p>PICO_OK^[23]</p> <p>PICO_INVALID_HANDLE^[23]</p> <p>PICO_NULL_PARAMETER^[23]</p> <p>PICO_INVALID_PARAMETER^[23]</p>

3.3.7 pl1000OpenUnit

```
PICO_STATUS pl1000OpenUnit(  
    short * handle  
)
```

This function opens and enumerates the unit.

Arguments:	handle: the function will write a value here that uniquely identifies the data logger that was opened. Use this as the <code>handle</code> parameter when calling any other PicoLog 1000 Series API function.
Returns:	PICO_OK ²³ PICO_OS_NOT_SUPPORTED ²³ PICO_OPEN_OPERATION_IN_PROGRESS ²³ PICO_EEPROM_CORRUPT ²³ PICO_KERNEL_DRIVER_TOO_OLD ²³ PICO_FW_FAIL ²³ PICO_MAX_UNITS_OPENED ²³ PICO_NOT_FOUND ²³ PICO_NOT_RESPONDING ²³

3.3.8 pl1000OpenUnitAsync

```
PICO_STATUS pl1000OpenUnitAsync(  
    short * status  
)
```

This function opens a PicoLog 1000 Series data logger without waiting for the operation to finish. You can find out when it has finished by periodically calling [pl1000OpenUnitProgress](#)^[15] until that function returns a non-zero value and a valid data logger handle.

The driver can support up to four data loggers.

Arguments:	status: the function writes a status flag to this location: 0 if there is already an open operation in progress 1 if the open operation is initiated
Returns:	PICO_OK ^[23] PICO_OPEN_OPERATION_IN_PROGRESS ^[23] PICO_OPERATION_FAILED ^[23]

3.3.9 pl1000OpenUnitProgress

```
PICO_STATUS pl1000OpenUnitProgress(
    short * handle,
    short * progress,
    short * complete
)
```

This function checks on the progress of [pl1000OpenUnitAsync](#)^[14].

Arguments:	<p>handle: a pointer to where the function should store the handle of the opened data logger, if the operation was successful. Use this as the <code>handle</code> parameter when calling any other PicoLog 1000 Series API function.</p> <p>0: if no unit is found or the unit fails to open <>0: handle of unit (valid only if function returns <code>PICO_OK</code>)</p> <p>progress: a location where the function writes an estimate of the progress towards opening the data logger. The value is between 0 to 100.</p> <p>complete: a location where the function will write a non-zero value if the operation has completed</p>
Returns:	<p>PICO_OK^[23]</p> <p>PICO_NULL_PARAMETER^[23]</p> <p>PICO_OPERATION_FAILED^[23]</p>

3.3.10 pl1000Ready

```
PICO_STATUS pl1000Ready(
    short    handle,
    short * ready
)
```

This function indicates when [pl1000Run](#)^[17] has captured the requested number of samples.

Arguments:	<p><code>handle</code>: handle returned from pl1000OpenUnit^[13] or pl1000OpenUnitProgress^[15]</p> <p><code>ready</code>: TRUE if ready, FALSE otherwise</p>
Returns:	<p>PICO_OK^[23]</p> <p>PICO_INVALID_HANDLE^[23]</p> <p>PICO_NOT_RESPONDING^[23]</p>

3.3.11 pl1000Run

```
PICO_STATUS pl1000Run(
    short      handle,
    unsigned long no_of_values,
    BLOCK_METHOD method
)
```

This function tells the unit to start capturing data.

Arguments:	<p>handle: handle returned from pl1000OpenUnit^[13] or pl1000OpenUnitProgress^[15]</p> <p>no_of_values: the number of samples the unit should collect</p> <p>method: which method to use to collect the data, from the following list:</p> <ul style="list-style-type: none"> BM_SINGLE BM_WINDOW BM_STREAM <p>See "Capture modes"^[6] for details.</p>
Returns:	<p>PICO_OK^[23]</p> <p>PICO_INVALID_HANDLE^[23]</p> <p>PICO_USER_CALLBACK^[23]</p> <p>PICO_INVALID_CHANNEL^[23]</p> <p>PICO_TOO_MANY_SAMPLES^[23]</p> <p>PICO_INVALID_TIMEBASE^[23]</p> <p>PICO_NOT_RESPONDING^[23]</p> <p>PICO_CONFIG_FAIL^[23]</p> <p>PICO_INVALID_PARAMETER^[23]</p> <p>PICO_NOT_RESPONDING^[23]</p> <p>PICO_TRIGGER_ERROR^[23]</p>

3.3.12 pl1000SetDo

```
PICO_STATUS pl1000SetDo(  
    short handle,  
    short do_value,  
    short doNo  
)
```

This function controls the digital outputs DO0 to DO3 on the unit.

Arguments:	<p><code>handle</code>: handle returned from pl1000OpenUnit¹³ or pl1000OpenUnitProgress¹⁵</p> <p><code>do_value</code>: whether to switch the output on or off: 1 - turns the digital output on 0 - turns the digital output off</p> <p><code>doNo</code>: which output to switch, from <code>PL1000_DO_CHANNEL_0</code> to <code>PL1000_DO_CHANNEL_3</code></p>
Returns:	<p>PICO_OK²³</p> <p>PICO_INVALID_HANDLE²³</p> <p>PICO_NOT_RESPONDING²³</p>

3.3.13 pl1000SetInterval

```
PICO_STATUS pl1000SetInterval(
    short        handle,
    unsigned long * us_for_block,
    unsigned long ideal_no_of_samples,
    short        * channels,
    short        no_of_channels
)
```

This function sets the sampling rate of the unit.

The fastest possible sampling interval is 1 microsecond, when the number of samples is 8129 divided by the number of channels active and the [capture mode](#)^[6] is `BM_SINGLE`. Under all other conditions, the fastest possible sampling interval is 10 microseconds.

Arguments:	<p><code>handle</code>: handle returned from pl1000OpenUnit^[13] or pl1000OpenUnitProgress^[15]</p> <p><code>us_for_block</code>: target total time in which to collect <code>ideal_no_of_samples</code>, in microseconds</p> <p><code>ideal_no_of_samples</code>: the number of samples that you want to collect. This number is used only for timing calculations.</p> <p><code>channels</code>: an array of numbers identifying the channels from which you wish to capture data, from <code>PL1000_CHANNEL_1</code> to <code>PL1000_CHANNEL_12</code> (for the PicoLog 1012) or <code>PL1000_CHANNEL_16</code> (for the PicoLog 1216).</p> <p><code>no_of_channels</code>: the number of channels in the <code>channels</code> array</p>
Returns:	<p>PICO_OK^[23]</p> <p>PICO_INVALID_HANDLE^[23]</p> <p>PICO_INVALID_CHANNEL^[23]</p> <p>PICO_INVALID_TIMEBASE^[23]</p> <p>PICO_NOT_RESPONDING^[23]</p> <p>PICO_CONFIG_FAIL^[23]</p> <p>PICO_INVALID_PARAMETER^[23]</p> <p>PICO_NOT_RESPONDING^[23]</p> <p>PICO_TRIGGER_ERROR^[23]</p>

3.3.14 pl1000SetPulseWidth

```
PICO_STATUS pl1000SetPulseWidth(
    short handle,
    unsigned short period,
    unsigned char cycle
)
```

This function sets the pulse width of the PWM (pulse-width modulated) output.

Arguments:	<p>handle: handle returned from pl1000OpenUnit^[13] or pl1000OpenUnitProgress^[15]</p> <p>period: the required period of the PWM waveform in microseconds, from 100 to 1800</p> <p>cycle: the required duty cycle as a percentage from 0 to 100</p>
Returns:	<p>PICO_OK^[23]</p> <p>PICO_INVALID_HANDLE^[23]</p> <p>PICO_SIG_GEN_PARAM^[23]</p> <p>PICO_NOT_RESPONDING^[23]</p>

3.3.15 pl1000SetTrigger

```
PICO_STATUS pl1000SetTrigger(
    short      handle,
    unsigned short enabled,
    unsigned short auto_trigger,
    unsigned short auto_ms,
    unsigned short channel,
    unsigned short dir,
    unsigned short threshold,
    unsigned short hysteresis,
    float      delay
)
```

This function sets up the trigger, which controls when the unit starts capturing data.

Arguments:	<p>handle: handle returned from pl1000OpenUnit^[13] or pl1000OpenUnitProgress^[15]</p> <p>enabled: whether to enable or disable the trigger: 0: disable the trigger 1: enable the trigger</p> <p>auto_trigger: whether or not to re-arm the trigger automatically after each trigger event: 0: do not auto-trigger 1: auto-trigger</p> <p>auto_ms: time in milliseconds after which the unit will auto-trigger if the trigger condition is not met</p> <p>channel: which channel to trigger on, from <code>PL1000_CHANNEL_1</code> to <code>PL1000_CHANNEL_12</code> (for the PicoLog 1012) or <code>PL1000_CHANNEL_16</code> (for the PicoLog 1216).</p> <p>dir: which edge to trigger on: 0: rising edge 1: falling edge</p> <p>threshold: trigger threshold (the level at which the trigger will activate) in ADC counts</p> <p>hysteresis: trigger hysteresis in ADC counts. This is the difference between the upper and lower thresholds. The signal must then pass through both thresholds in the same direction in order to activate the trigger, so that there are fewer unwanted trigger events caused by noise. The minimum value allowed is 1.</p> <p>delay: delay between the trigger event and the start of the block as a percentage of the block size. 0% means that the trigger event is the first data value in the block, and -50% means that the trigger event is in the middle of the block.</p>
Returns:	<p>PICO_OK^[23] PICO_INVALID_HANDLE^[23] PICO_USER_CALLBACK^[23] PICO_TRIGGER_ERROR^[23] PICO_MEMORY_FAIL^[23]</p>

3.3.16 pl1000Stop

```
PICO_STATUS pl1000Stop(  
    short handle  
)
```

This function aborts data collection.

Arguments:	<code>handle</code> : handle returned from pl1000OpenUnit ¹³ or pl1000OpenUnitProgress ¹⁵
Returns:	PICO_OK ²³ PICO_INVALID_HANDLE ²³

3.3.17 PICO_STATUS values

Every function in the PicoLog 1000 Series API returns an error code from the following list of PICO_STATUS values:

Code (hex)	Enum	Description
00	PICO_OK	The Data Logger is functioning correctly
01	PICO_MAX_UNITS_OPENED	An attempt has been made to open more than <code>PL1000_MAX_UNITS</code>
02	PICO_MEMORY_FAIL	Not enough memory could be allocated on the host machine
03	PICO_NOT_FOUND	No PicoLog 1000 device could be found
04	PICO_FW_FAIL	Unable to download firmware
05	PICO_OPEN_OPERATION_IN_PROGRESS	A request to open a device is in progress
06	PICO_OPERATION_FAILED	The operation was unsuccessful
07	PICO_NOT_RESPONDING	The device is not responding to commands from the PC
08	PICO_CONFIG_FAIL	The configuration information in the device has become corrupt or is missing
09	PICO_KERNEL_DRIVER_TOO_OLD	The <code>picopp.sys</code> file is too old to be used with the device driver
0A	PICO_EEPROM_CORRUPT	The EEPROM has become corrupt, so the device will use a default setting
0B	PICO_OS_NOT_SUPPORTED	The operating system on the PC is not supported by this driver
0C	PICO_INVALID_HANDLE	There is no device with the handle value passed
0D	PICO_INVALID_PARAMETER	A parameter value is not valid
0E	PICO_INVALID_TIMEBASE	The time base is not supported or is invalid
0F	PICO_INVALID_VOLTAGE_RANGE	The voltage range is not supported or is invalid
10	PICO_INVALID_CHANNEL	The channel number is not valid on this device or no channels have been set
11	PICO_INVALID_TRIGGER_CHANNEL	The channel set for a trigger is not available on this device
12	PICO_INVALID_CONDITION_CHANNEL	The channel set for a condition is not available on this device
13	PICO_NO_SIGNAL_GENERATOR	The device does not have a signal generator
14	PICO_STREAMING_FAILED	Streaming has failed to start or has stopped without user request
15	PICO_BLOCK_MODE_FAILED	Block failed to start - a parameter may have been set wrongly
16	PICO_NULL_PARAMETER	A parameter that was required is <code>NULL</code>
18	PICO_DATA_NOT_AVAILABLE	No data is available from a run block call
19	PICO_STRING_BUFFER_TOO_SMALL	The buffer passed for the information was too small
1A	PICO_ETS_NOT_SUPPORTED	ETS is not supported on this device
1B	PICO_AUTO_TRIGGER_TIME_TOO_SHORT	The auto trigger time is less than the time it will take to collect the data
1C	PICO_BUFFER_STALL	The collection of data has stalled as unread data would be overwritten
1D	PICO_TOO_MANY_SAMPLES	The number of samples requested is more than available in the current memory segment
1E	PICO_TOO_MANY_SEGMENTS	Not possible to create number of segments requested
1F	PICO_PULSE_WIDTH_QUALIFIER	A null pointer has been passed in the trigger function or one of the parameters is out of range

20	PICO_DELAY	One or more of the hold-off parameters are out of range
21	PICO_SOURCE_DETAILS	One or more of the source details are incorrect
22	PICO_CONDITIONS	One or more of the conditions are incorrect
23	PICO_USER_CALLBACK	The driver's thread is currently in the pl1000Ready ^[16] callback function and therefore the action cannot be carried out
24	PICO_DEVICE_SAMPLING	An attempt is being made to get stored data while streaming. Either stop streaming by calling pl1000Stop , ^[22] or use ps4000GetStreamingLatestValues ^[11]
25	PICO_NO_SAMPLES_AVAILABLE	...because a run has not been completed
26	PICO_SEGMENT_OUT_OF_RANGE	The memory index is out of range
27	PICO_BUSY	Data cannot be returned yet
28	PICO_STARTINDEX_INVALID	The start time to get stored data is out of range
29	PICO_INVALID_INFO	The information number requested is not a valid number
2A	PICO_INFO_UNAVAILABLE	The handle is invalid so no information is available about the device. Only <code>PICO_DRIVER_VERSION</code> is available.
2B	PICO_INVALID_SAMPLE_INTERVAL	The sample interval selected for streaming is out of range
2C	PICO_TRIGGER_ERROR	Not used
2D	PICO_MEMORY	Driver cannot allocate memory
36	PICO_DELAY_NULL	<code>NULL</code> pointer passed as delay parameter
37	PICO_INVALID_BUFFER	The buffers for overview data have not been set while streaming
3A	PICO_CANCELLED	A block collection has been cancelled
3B	PICO_SEGMENT_NOT_USED	The segment index is not currently being used
3C	PICO_INVALID_CALL	The wrong GetValues ^[11] function has been called for the collection mode in use
3F	PICO_NOT_USED	The function is not available
41	PICO_INVALID_STATE	Device is in an invalid state
43	PICO_DRIVE_FUNCTION	You called a driver function while another driver function was still being processed

3.4 Example programs

3.4.1 Introduction

We supply examples for the following programming languages:

- [C](#) ²⁵ and [C++](#) ²⁵
- [Delphi](#) ²⁵
- [Excel](#) ²⁵
- [LabVIEW](#) ²⁶
- [Visual Basic](#) ²⁶
- [Agilent VEE](#) ²⁶

3.4.2 C and C++

C

Use the following files:

```
pl1000.lib  
pl1000api.h  
pl1000bc.lib  
pl1000con.c
```

C++

C++ programs can access all versions of the driver. If `pl1000api.h` is included in a C++ program, the `PREF1` macro expands to `extern "C"`: this disables name-decoration, and enables C++ routines to make calls to the driver routines using C headers.

3.4.3 Delphi

`pl1000.dpr` is a complete program that opens the driver and reads values from Channel 1.

The file `pl1000.inc` contains a set of procedure prototypes that you can include in your own programs.

You will also require the following files:

```
pl1000.vh  
pl1000fm.dfm  
pl1000.pas
```

3.4.4 Excel

The easiest way to transfer data to Excel is to use PicoLog. However, you can also write an Excel macro which calls `pl1000.dll` to read in a set of data values. The Excel macro language is similar to Visual Basic.

The example `pl1000.xls` reads in 20 values from Channels 1 and 2, one per second, and assigns them to cells A1..B20.

Use Excel Version 7 or higher.

Note that it is usually necessary to copy the DLL file to the `\windows\system` directory.

3.4.5 LabVIEW

The routines described here were tested using LabVIEW version 4.0.

While it is possible to access all of the driver routines described earlier, it is easier to use the special LabVIEW access routines if only single readings are required. The `pl1000.lib` library in the installation directory shows how to access these routines.

To use these routines, copy `pl1000.lib` and `pl1000.dll` to your LabVIEW `user.lib` directory. You will then find a sub-vi to access the logger, and some example sub-vis which demonstrate how to use it.

You can use one of these sub-vis for each of the channels that you wish to measure. The sub-vi accepts the port and channel (1 to 11) and returns a voltage.

3.4.6 Visual Basic

Version 4 and 5 (32-bit)

The installation directory contains the following files:

```
pl100032.VBP  
pl100032.BAS  
pl100032.FRM
```

3.4.7 Agilent-VEE

The example routine `pl1000.vee` is in the installation directory. It was tested using Agilent VEE version 5. The example shows how to collect a block of data from the ADC.


4 Glossary

ADC. Analog to Digital Converter. An ADC samples analog signals and converts them to digital data for storage and processing. It is an essential component of a data logger.

DLL. Dynamic Link Library. A file containing a collection of Windows functions designed to perform a specific class of operations. A DLL is supplied with the PicoLog Data Loggers to enable you to control the devices from your own programs.

Driver. A small program that acts as an interface, generally between a hardware component and a computer program. The PicoLog Data Loggers require a USB driver that runs in the Windows kernel, and a second driver in the form of a DLL that communicates with your application.

Maximum sampling rate. A figure indicating the maximum number of samples the ADC is capable of acquiring per second. Maximum sample rates are usually given in S/s (samples per second). The higher the sampling rate of the ADC, the more accurately it can represent the high-frequency details in a signal.

Streaming. An operating mode in which the [ADC](#)  samples data and returns it to the computer in an unbroken stream.

USB. Universal Serial Bus. This is a standard port that enables you to connect external devices to PCs. A typical USB 1.1 port supports a data transfer rate of 12 megabits per second and is much faster than an RS-232 serial port.

Index

6

64-bit Windows 4

A

ADC value, maximum 12
 ADC-11 compatibility 5
 Agilent VEE 26
 Asynchronous operation 6

B

BM_SINGLE mode 6
 BM_STREAM mode 6
 BM_WINDOW mode 6

C

C 25
 C++ 25
 Capture modes
 BM_SINGLE 6
 BM_STREAM 6
 BM_WINDOW 6
 Closing a unit 8
 Connecting to the PC 4
 Contact details 3

D

Data, reading 9, 11
 Delphi 25
 Digital outputs, setting 18
 DLLs 4
 Driver routines
 pl1000CloseUnit 8
 pl1000GetSingle 9
 pl1000GetUnitInfo 10
 pl1000GetValues 11
 pl1000MaxValue 12
 pl1000OpenUnit 13
 pl1000OpenUnitAsync 14
 pl1000OpenUnitProgress 15
 pl1000Ready 16
 pl1000Run 17
 pl1000SetDo 18
 pl1000SetInterval 19
 pl1000SetPulseWidth 20
 pl1000SetTrigger 21

pl1000Stop 22
 summary 7

E

Excel 25

G

Glossary 27

I

Information on unit, obtaining 10
 Installation 4

L

LabVIEW 26
 Legal information 2

M

Maximum ADC value 12

N

New Hardware Wizard 4

O

Opening a unit 13, 14, 15, 16
 Overview 1

P

PicoLog 100 Series SDK 5
 Programming 4, 25
 Programming languages
 Agilent VEE 26
 C 25
 C++ 25
 Delphi 25
 Excel macros 25
 LabVIEW 26
 Visual Basic 26
 Pulse width, setting 20
 PWM output, setting up 20

R

Running a unit 17

S

- Sampling interval, setting 19
- Scaling 6
- SDK 5
- Stopping a unit 22
- Streaming 6

T

- Trigger, setting 21

U

- Unit information, obtaining 10
- USB ADC-11 compatibility 5

V

- Visual Basic 26

W

- Windows XP/Vista
 - support 4
- WoW64 4



Pico Technology

James House
Colmworth Business Park
ST. NEOTS
Cambridgeshire
PE19 8YP
United Kingdom
Tel: +44 (0) 1480 396 395
Fax: +44 (0) 1480 396 296
www.picotech.com

pl1000pg.en-1

13.5.09

Copyright © 2009 Pico Technology Ltd. All rights reserved.